

## GF3000-G

### EMC ANTI-INTERFERENCE METERING TEST SYSTEM

GF3000-G EMC anti-interference metering test system is a high-end equipment designed for electricity meter electromagnetic compatibility (EMC) test developed and produced by GFUVE GROUP. The products are equipped with EMC filter purification device, GF303G three-phase AC power source, GF333G three-phase AC reference meter, multi-channel digital information transmission server, Beidou precision time base source system and PC software test platform developed and produced by our company, which are controlled and managed by computer. The device can simulate the complex electromagnetic interference environment of electrostatic discharge (ESD), Electrical Fast Transient (EFT), RF field conduction/radiation immunity, lightning surge and other complex electromagnetic interference environments, and comprehensively verify the anti-interference ability and stability of electricity meters under strict working conditions, in line with the latest applicable IEC standards and corresponding reasonable standards. It is strictly in accordance with IEC60736, IR46 and ANSI C12.20 standard, meeting ISO17025 laboratory standards!

## Functions

1. Timing error test.
2. Creeping and starting test.
3. Standard deviation, 24 hour variation test.
4. Equipped with RS232 communication port.
5. With PC auto test energy meter accuracy software.
6. Calibration and verification of energy meter constant.
7. Influencing quantity test of voltage, frequency, harmonic.
8. Able to check the same type meters with different constants.
9. Full automatic, semi automatic and manual operation available.
10. Supporting bar code input, which increase the efficiency of input.
11. Able to print out all kinds of test reports with the standard forms.
12. Testing energy registers (dial test) and maximum demand indicator.
13. Forward and reverse active energy / forward and reverse reactive energy error test.
14. Able to calibrate all kinds of electronic and inductive three phase kWh meter in EMC laboratory.
15. Able to perform automatic measurement like shunting, basic errors, standard deviation, and etc.
16. Able to test voltage, current, active/reactive/apparent power, phase, power factor, frequency, and etc.
17. Able to display waveform of voltage and current able to set 2nd-63st harmonic of voltage and current, measure the waveform distortion and harmonic content, and display harmonic chart.



## Features

1. Operate automatically or manually;
2. With GPS BEIDOU Precision time server;
3. High resolution of voltage, current and power;
4. Large LCD display and simple interface for operation;
5. Power source have DIP and interrupts setting;
6. Three phase reference meter and three phase power source with touch TFT color LCD;
7. High accuracy, 6-digit display the energy relative errors are no more than 0.05% within the measuring range;
8. High stability of power source which is up to 0.01%/100s and low distortion which is no more than 0.3%;
9. Wide current measuring range from 1mA to 120A which can be automatically switched;
10. Self-check and perfect protection function of overload, short voltage circuit and open current circuit;
11. Equipped with three phase multifunction reference meter and program-controlled three phase power source which can be separately used and are convenient for testing;

## Parameters

Electrical parameters	
Accuracy	0.02%, 0.05%, 0.1%
Power Supply	AC 180-265V, or 3×220/380V±15%, frequency 50/60Hz.
Three Phase Reference Meter	
Voltage measurement	
Range(U1, U2, U3)	0.00-600V
Resolution	0.001V
Error	±0.02% (30V-600V) ±0.05% (0.1V-30V)
Harmonic	2 <sup>nd</sup> -63 <sup>rd</sup>
Current measurement	
Range(I1,I2,I3)	1mA-120A
Resolution	0.1mA
Error	±0.02% (10mA-120A) ±0.05% (1mA-10mA)
Harmonic	2 <sup>nd</sup> -63 <sup>rd</sup>
Power measure error	
Active power (direct connection)	±0.02% (0.01A-120A) or ±0.05% (0.01A-120A) ±0.05% (0.001A-0.01A) or ±0.1% (0.001A-0.01A)
Reactive power (direct connection)	±0.05% (0.1A-120A) or ±0.1% (0.1A-120A)

### Electrical parameters - continued

#### Phase angle measurement

Range	0°-360°
Resolution	0.005°
Error	±0.015°

#### Frequency Measurement

Range	40-70Hz
Resolution	0.0005Hz
Error	0.001Hz

#### Power factor measurement

Range	-1.0 ~ 0 ~ +1.0
Resolution	0.0001
Error	0.0005

#### Three Phase Power Source

##### AC Voltage Output

Range(U1, U2, U3)	57.7V, 100V, 220V, 380V (max 480V)
Adjustment range	(0-120)%RG <sup>(1)</sup>
Adjustment fineness	0.01%RG, 0.1%RG, 1%RG, 10%RG as optional.
Stability	0.01%/120s
Distortion	0.08% (Non-capacitive load)
Output load	1200VA/Phase
Measuring accuracy	0.02%RG or 0.05% RG

##### AC Current Output

Range(I1,I2,I3)	0.01A, 0.05A, 0.1A, 0.5A, 1A, 5A, 10A, 50A, 100A
Adjustment range	(0-120)%RG
Adjustment fineness	0.01%RG, 0.1%RG, 1%RG, 10%RG as optional.
Stability	<0.01%/120s
Distortion	≤0.08% (Non-capacitive load)
Output load	2400VA/Phase
Accuracy	0.02%RG or 0.05% RG

##### Power Output

Active power output stability	<0.01%RG/120s
Reactive power output stability	<0.02%RG/120s
Active power measuring accuracy	0.02%RG or 0.05% RG
Reactive power measuring accuracy	0.1%RG

##### Phase Output

Output adjustment range	0°-359.999°
Output adjustment fineness	10, 1, 0.1, 0.01 as optional.
Resolution	0.01°
Accuracy	0.02° or 0.05°

### Electrical parameters - continued

#### Power Factor

Adjustment range	-1 ~ 0 ~ 1
Resolution	0.0001
Measurement accuracy	0.0005

#### Frequency Output

Adjustment range	40Hz-70Hz
Output adjustment fineness	5Hz, 1Hz, 0.1Hz, 0.01Hz as optional.
Resolution	0.001Hz
Accuracy	0.002Hz

#### Voltage /Current/Harmonic Setting

Harmonic number	2-63 times
Harmonic content	0-40%
Harmonic phase	0-359.99
Harmonic setting accuracy	(10%±0.1%)RD <sup>(2)</sup>

#### Power Energy Measurement Error

Active power energy	0.02%RG or 0.05% RG
Reactive power energy	0.1%RG

#### Power Pulse Output

Power pulse type	active pulse, reactive pulse
Active power pulse output	5V, 10mA
Pulse output frequency	Max 50kHz

#### Power Pulse Input

Pulse constant set range	(1-599999999)/kwh
Energy pulse type	support active and reactive pulse, the highest frequency power pulse input is 200KHz.

#### Meter Position

Position	one meter
----------	-----------

#### Standard

Standard	IEC 62053-21,22, 23; IEC 60736; ANSI C12.20-2002; JJG 597-2005; JJG596-2012; JJG 1085-2013; JJF 68-2019; DL/T 826-2002; DL/T 1478-2015; DL/T 448-2016; EN 50470-1, EN 50470-2, EN-50470-3; IEC 61010;
----------	---

#### Safety

Isolation protection	IEC 61010-1:2001
Measurement Category	300 V CAT III, 600 V CAT II
Degree of protection	IP20
Declaration of conformity	CE & CNAS certified

#### Mechanical parameters

Dimensions (mm)	Cabinet size: 800 * 600 * 1850mm (L * W * H).
Weight (kg)	About 500

### Environmental conditions

Ambient temperature	0°C to +40°C
Relative humidity	35%-85%


(1) RG means range, the same as below;

(2) RD means the setted harmonic content, harmonic can be a single output, also multiple output.

## GF3000-20 Posion Packing List

Item	Name	QTY
1	GF333G three phase reference meter	1PC
2	GF303G three phase power source	1PC
3	BEIDOU precision time base source	1PC
4	EMC filter purification device	1PC
5	Scanning head	1PC
6	Energy pulse cable(input)	1PC
7	Auxiliary test cable 10m	1SET
8	PC computer with automatic testing software	1PC
9	Operation Manual	1PC
10	Factory test certificate	1PC

## PC automatic Testing Software (Optional)



The screenshot shows the Electric Energy Meter Verification software interface. The top menu bar includes: File, ParameterEntry, ResultSummary, DetailResult, DataSave, Online, PowerOn, PowerOff, Schema, and System. The main window has a toolbar with icons for ParameterEntry, ResultSummary, DetailResult, DataSave, Online, PowerOn, PowerOff, Schema, and System. The ResultSummary tab is active, displaying a table of test results for AccuracyRequirementTest. The table has columns for AccuracyRequirementTest, BasicError, and various parameters like No\_P+, H\_Imax\_1.0, No\_P+, H\_Imax\_0.5L, etc. The RunLog tab at the bottom shows a log of events from 2025-06-16 09:38:29 to 2025-06-16 09:38:33, including device port parameters loaded and verification control service connected. The bottom status bar shows: three-phase machine 08390, Stop Verification, Continuous mode BasicErrorTest (1/80) BasicError\_No\_P+\_H\_Imax\_1.0, and Verifier Admin.



Electric Energy Meter Verification Version : 1.0.13.2013

ParameterEntry ResultSummary DetailResult DataSave Online PowerOn PowerOff Schema System

ResultSummary DataSave

Detailed Conclusions ResultSummary

Table	Check	Factory No.	Conclusion	TestTime	PowerDevice	PowerDirection	CurrentRatio	PowerFactor	LowerLimit	UpperLimit	ErrorLaps	Error1	Error2	Avg	
1	<input checked="" type="checkbox"/>	12	Qualified	BasicError_No_P+_H_Imax_1.0	2022-12-12 16:48:59	H	正向有功	Imax	1.0	-1	1	24	-0.0543	-0.0735	-0.0644
2	<input type="checkbox"/>	13	Qualified	BasicError_No_P+_H_Imax_0.5L	2022-12-12 16:52:16	H	正向有功	Imax	0.5L	-1	1	12	-0.1998	-0.2353	-0.2125
3	<input type="checkbox"/>	125	Qualified	BasicError_No_P+_H_Imax_0.8C	2022-12-12 16:52:58	H	正向有功	Imax	0.8C	-1	1	19	+0.0492	+0.0350	+0.0421
4	<input type="checkbox"/>	126	Qualified	BasicError_No_P+_H_0.5Imax_1.0	2022-12-12 16:53:40	H	正向有功	0.5Imax	1.0	-1	1	12	+0.0209	+0.0212	+0.0210
5	<input type="checkbox"/>	127	Qualified	BasicError_No_P+_H_0.5Imax_0.5L	2022-12-12 16:54:23	H	正向有功	0.5Imax	0.5L	-1	1	6	-0.1562	-0.1400	-0.1481
6	<input type="checkbox"/>	128	Qualified	BasicError_No_P+_H_0.5Imax_0.8C	2022-12-12 16:55:05	H	正向有功	0.5Imax	0.8C	-1	1	10	+0.0900	+0.0961	+0.0930
7	<input type="checkbox"/>	129	Qualified	BasicError_No_P+_H_1b_1.0	2022-12-12 16:55:51	H	正向有功	1b	1.0	-1	1	2	+0.0598	+0.0599	+0.0598
8	<input type="checkbox"/>	130	Qualified	BasicError_No_P+_H_1b_0.5L	2022-12-12 16:56:33	H	正向有功	1b	0.5L	-1	1	1	+0.0342	+0.0049	+0.0200
9	<input type="checkbox"/>	131	Qualified	BasicError_No_P+_H_1b_0.8C	2022-12-12 16:57:19	H	正向有功	1b	0.8C	-1	1	2	+0.1061	+0.0967	+0.0964
10	<input type="checkbox"/>	132	Qualified	BasicError_No_P+_H_0.1lb_1.0	2022-12-12 16:58:05	H	正向有功	0.1lb	1.0	-1	1	1	+0.0675	+0.0769	+0.0722
11	<input type="checkbox"/>	133	Qualified	BasicError_No_P+_H_0.1lb_0.5L	2022-12-12 16:59:31	H	正向有功	0.1lb	0.5L	-1.5	1.5	1	+0.0655	+0.0776	+0.0715
12	<input type="checkbox"/>	134	Qualified	BasicError_No_P+_H_0.1lb_0.8C	2022-12-12 17:02:02	H	正向有功	0.1lb	0.8C	-1.5	1.5	1	+0.0772	+0.0828	+0.0800
				BasicError_No_P+_H_0.03lb_1.0	2022-12-12 17:03:46	H	正向有功	0.03lb	1.0	-1.5	1.5	1	+0.0337	+0.0556	+0.0446
				BasicError_No_P+_A_Imax_1.0	2022-12-12 17:08:27	A	正向有功	Imax	1.0	-2	2	8	+0.0563	+0.0592	+0.0577
				BasicError_No_P+_A_Imax_1.0	2022-12-12 17:08:27	A	正向有功	Imax	1.0	-2	2	8	+0.0563	+0.0592	+0.0577
				BasicError_No_P+_A_Imax_0.5L	2022-12-12 17:09:55	A	正向有功	Imax	0.5L	-2	2	4	-0.2529	-0.2281	-0.2355
				BasicError_No_P+_A_Imax_0.5L	2022-12-12 17:09:55	A	正向有功	Imax	0.5L	-2	2	4	-0.2529	-0.2281	-0.2355
				BasicError_No_P+_A_Imax_0.8C	2022-12-12 17:11:23	A	正向有功	Imax	0.8C	-2	2	6	+0.1912	+0.1630	+0.1771
				BasicError_No_P+_A_Imax_0.8C	2022-12-12 17:11:23	A	正向有功	Imax	0.8C	-2	2	6	+0.1912	+0.1630	+0.1771
				BasicError_No_P+_A_1b_1.0	2022-12-12 17:12:54	A	正向有功	1b	1.0	-2	2	1	+0.1638	+0.1699	+0.1663
				BasicError_No_P+_A_1b_1.0	2022-12-12 17:12:54	A	正向有功	1b	1.0	-2	2	1	+0.1638	+0.1699	+0.1663
				BasicError_No_P+_A_1b_0.5L	2022-12-12 17:14:59	A	正向有功	1b	0.5L	-2	2	1			
				BasicError_No_P+_A_1b_0.5L	2022-12-12 17:14:59	A	正向有功	1b	0.5L	-2	2	1			
				BasicError_No_P+_A_1b_0.8C	2022-12-12 17:21:02	A	正向有功	1b	0.8C	-2	2	1	+0.2093	+0.2211	+0.2152

RunLog

2025-06-16 09:38:29 DeviceOperationLogs The device port parameters are successfully loaded.

2025-06-16 09:38:29 DeviceOperationLogs Start issuing parameters and establishing device connection, please wait

2025-06-16 09:38:29 VerificationBusinessLog The verification message service was created successfully. The address is: http://127.0.0.1:8888

2025-06-16 09:38:29 VerificationBusinessLog The local verification conclusion service is created successfully. The address is: http://127.0.0.1:8889

2025-06-16 09:38:32 VerificationBusinessLog Start connecting to the verification control service at: http://127.0.0.1:12121

2025-06-16 09:38:33 VerificationBusinessLog The connection to the verification control service is completed, and the verification control client has been created

2025-06-16 09:38:33 VerificationBusinessLog The verification control authority was obtained successfully!

2025-06-16 09:38:33 VerificationBusinessLog Issued by the Execution Machine Configuration Information

Standard Meter x RunLog x

three-phase machine 08390 Stop Verification Continuous mode BasicErrorTest (1/80) BasicError\_No\_P+\_H\_Imax\_1.0

Verifier Admin Auditor Admin Supervisor